

# F668 Ophthalmology Electric Chair (new product)

## F668 Specification :

1. The seat is made of fiberglass base and paint material.
2. The Cushion is made of high-density foam and ventilate fur.
3. Seat Lying Length: 185cm.
4. The seat width: 43cm.
5. Containing armrest seat 57cm width
6. Chair Height: High: 72 cm Low: 49 cm.
7. The seat lifter up-down range: 23cm.
8. Rotation range: manually rotated 180 degrees.
9. Back tilt and lift height controlled and adjusted to the desired angle by electric.
10. The back inverted 180 degrees by electric, the pedal can be hence raised horizontally. Armrest can be downward to the level of seat.
11. Headrest: angle formed adjustable by lever and fixed.
12. Optional item: U-type fixed pillow.
13. The device automatically recovery: seat can be automatically downward.
14. Voltage: AC220V Current: 5A
15. The style can be modified per customer's request.

